

19104268D

1 SENATE BILL NO. 1760

2 Offered January 18, 2019

3 A BILL to amend and reenact § 32.1-229.1 of the Code of Virginia, relating to diagnostic X-ray
4 machines; operation.

5 Patron—DeSteph

6 Referred to Committee on Education and Health

7 Be it enacted by the General Assembly of Virginia:

8 1. That § 32.1-229.1 of the Code of Virginia is amended and reenacted as follows:

9 § 32.1-229.1. Inspections of X-ray machines required; Radiation Inspection Reports; fees;
10 qualification of inspectors.

11 A. All X-ray machines shall be registered with the Department.

12 B. Every owner or operator of an X-ray machine shall request an initial inspection by a private
13 inspector or a Department inspector no later than 30 days after the installation of the equipment.14 Inspections shall be performed periodically on a schedule prescribed by the Board. The Department
15 may also require random, unannounced, follow-up inspections of machines that were inspected by
16 private inspectors in order to maintain quality control. In the event of changes in or installations of new
17 equipment during the last 90 days of a period for which an inspection has been made, no interim
18 inspection shall be required. In addition, the Department may require the inspection and certification of
19 other machines emitting radiation or utilizing radiation for patients, consumers, workers, or the general
20 public.21 Inspections shall be performed by Department personnel or by private inspectors only. Inspections
22 conducted by private inspectors shall be conducted in conformance with the regulations of the Board
23 and reports on these inspections shall be filed by the registrant with the Department on forms prescribed
24 by the Department. Results of all inspections shall be reviewed by the Department.25 C. The Department shall issue a certificate for a diagnostic or therapeutic X-ray machine, or X-ray
26 machine not used in the healing arts, when the results of the inspection indicate the machine meets the
27 Board's standards. If the machine does not meet the Board's standards, the certification may be denied.
28 If the certification is denied, the machine shall not be used for treatment, diagnosis, evaluation of
29 patients, whether human or animal, or any other use until the standards of the Board have been met. A
30 copy of the certificate shall be displayed by the registrant in a conspicuous place in close proximity to
31 the X-ray machine.32 D. The Board shall, in accordance with the Administrative Process Act (§ 2.2-4000 et seq.),
33 promulgate such regulations as the Board deems necessary to protect the health and safety of health care
34 workers, patients, and the general public, including but not limited to:

35 1. Fee schedules for registration of X-ray machines;

36 2. Schedule for inspections of X-ray machines;

37 3. Fee schedules for inspections of X-ray machines by Department personnel; however, no fee shall
38 be charged for inspections initiated by the Department;

39 4. Standards for certification of X-ray machines; and

40 5. Qualifications for private inspectors of X-ray machines required for inclusion on a list of qualified
41 inspectors of X-ray machines published pursuant to § 32.1-228.1, a requirement for annual registration as
42 a private inspector of X-ray machines for inclusion on such list, and a fee not to exceed \$150.00 for
43 such registration.44 E. No person who has been trained and certified in the operation of a diagnostic X-ray machine by
45 the manufacturer of such machine shall be required to obtain any other training, certification, or
46 licensure or be under the supervision of a person who has obtained training, certification, or licensure
47 to operate such a diagnostic X-ray machine, provided that (i) the diagnostic X-ray machine (a) is
48 registered and certified by the Department, (b) is being operated to conduct a body composition scan,
49 and (c) is not operated to determine bone density or in the diagnosis or treatment of a patient and (ii)
50 the subject of the body composition scan is notified of the risks associated with exposure to radiation
51 emitted by the diagnostic X-ray machine.52 F. The provisions of this section and of §§ 32.1-229 and 32.1-229.2 relating to X-ray machines and
53 machines emitting or utilizing radiation shall not apply to devices purchased or used primarily for
54 personal, family, or household purposes.

INTRODUCED

SB1760